

## COMPASS cabling with amp-rack No.1 or 2

Connect the cabling as indicated in the figure below.

**Hint:** Keep connected **Ext-gradient cable** since the device monitoring via drive console requires the feedback.

**Hint:** For using the amp-rack with the COMPASS unit, switch on ALL amplifiers although only CH5 is used for field generation.

